

# GAS PRESSURE SENSOR W24

## USER GUIDE



[cma-science.nl](http://cma-science.nl)

## Short description

CMA Wireless Gas Pressure sensor W24 is designed to measure gas pressure, relative to ambient air pressure<sup>1</sup>, in the range between -100 to 300 kPa. The pressure is measured via a pressure port located on the side of the sensor box. The end of the white pressure port has a small threaded end called a Luer-lock. With a gentle half-turn, you may directly attach the plastic Luer-lock connector or the 20 mL plastic syringe (both included with the sensor).

The Gas Pressure sensor is a very precise piezoresistive pressure sensor. It consists of a silicon diaphragm with piezoresistive strain gauges diffused into it, fused to a glass backplate. Applied pressure causes mechanical stress of the diaphragm, which is transduced into a resistance change using the piezoresistive effect.

The sensor is fairly durable, but it is designed only for use with non-corrosive, non-ionic working gasses such as air, dry gases and the like. Do not get it wet.

The sensor is delivered with the following accessories:

- one plastic 20-ml syringe with Luer-lock,
- one plastic Luer-lock connector, and
- one silicone tube with an inside diameter of circa 6 mm. This tube can be attached to the other end of the Luer-lock connector.

The power button located on the top of the sensor allows you to turn the sensor on and off. The sensor is equipped with an OLED color display which shows some sensor information and the measured by the sensor values. This makes the sensor suitable to use as an independent measuring instrument.

The sensor can be used wirelessly via Bluetooth or wired via USB with the Coach 7 or Coach 7 Lite programs/apps on computers (Windows and Mac), Chromebooks and mobile devices (Android and iOS).

## Calibration

The Gas Pressure sensor W24 converts measured pressure values to digital values. It uses 14-bit analogue-to-digital conversion resulting in a resolution of 0.06 kPa. The sensor is supplied with a factory calibration in kPa.

When working with the Coach program the pre-defined calibration can be shifted by using the **Set to Value** option.

---

<sup>1</sup> Zero-referenced against ambient air pressure, so it is equal to absolute pressure minus atmospheric pressure.

## Software

You can use the Gas Pressure sensor W24 with Coach 7 or Coach 7 Lite (free) program on computers (Windows and Mac) or Coach 7 and Coach 7 Lite (free) app on mobile devices (Android and iOS). For Chromebooks, we offer a special Android app.

The support for wireless sensors is added starting from Coach version 7.10.



Check the CMA website for the latest installations.

[https://cma-science.nl/downloads\\_en](https://cma-science.nl/downloads_en)

## Collecting data without software connection

- Turn on the Gas Pressure sensor by pressing its power button.
- The sensor briefly displays its Bluetooth identification code. This code is also printed on the sticker located on the bottom side of the sensor box.
- Then the display shows:
  - the Bluetooth mode, 'Mobile' or 'PC'.  
Mobile indicates Bluetooth Low Energy mode which should be used when working with mobile devices (Android, iOS), Chromebook and Apple computers.  
PC indicates Bluetooth Classic which should be used for Windows computers.
  - the battery level, and
  - the current measured value.
- Now you can use the sensor as an independent measuring instrument.
- To turn off the sensor press and hold its power button for 3 sec. To save its battery the sensor automatically turns off after 5 minutes of inactivity (no connection to power, no communication).

## Collecting data via the Bluetooth connection

### ***Mobile devices, Chromebooks, and Apple computers***

For mobile devices (Android, iOS), Chromebooks and Apple computers Bluetooth Low Energy technology is used for wireless communication. For these devices **do not pair** the sensor just use it directly in the Coach software.

- Turn the Gas Pressure sensor on.
- Ensure your sensor is set to Mobile mode.  
If the display shows in the top-left corner 'PC' first you must set the sensor to the Mobile mode. Turn off the sensor. Then press and hold the power button until the text 'Bluetooth mode Change Mobile' is shown, then release the button. The mode is set to 'Mobile' which means that Bluetooth Low Energy is used.
- Start the Coach 7 or Coach 7 Lite program/app.
- Select the Dashboard Activity 'Measurement with Wireless sensors'.
- On opening of the Activity Coach starts searching for sensors which are turned on and in the Mobile discovery mode. The found Bluetooth sensors appear in the

list.

- Select the Gas Pressure sensor you want to connect to. If needed check the sensor's Bluetooth ID which is located on the sensor's bottom label.
- When the connection is established the Bluetooth symbol appears in the top-left corner of the sensor's display and the icon of the sensor appears showing the measured values.
- Now you are ready to use the Gas Pressure sensor for your measurement.

### **Windows computers**

For Windows computers, Bluetooth Classic technology is used for wireless communication. Before you start to use the sensor for measurement in Coach you have to **pair** it.

- Turn the Gas Pressure sensor on.
- Ensure your sensor is set to PC mode.  
If the display shows in the top-left corner 'Mobile' first you must set the sensor to the PC mode. Turn off the sensor. Then press and hold the power button until the text 'Bluetooth mode Change PC' is shown, then release the button. The mode is set to 'PC', meaning Bluetooth Classic is used.
- Pair your sensor.
  - Go to the Windows Settings **Bluetooth and other devices** and select **Add Bluetooth or other devices**. Select **Bluetooth device**.
  - Windows looks for Bluetooth devices and after a while lists discovered devices. The wireless sensors are listed with their Bluetooth IDs.
  - Select the sensor you want to connect to. If needed check the sensor's Bluetooth ID which is located on the bottom label of your sensors.
  - When the connection is successfully established Windows indicates that the sensor is paired and ready to go.
  - Click **Done** to accept it. The sensor appears in the list of paired Bluetooth devices.
- Start the Coach 7 or Coach 7 Lite program.
- Select the Dashboard Activity 'Measurement with Wireless sensors'.
- Coach starts searching and displays the list with detected sensors, even if they are not paired.
- Select the Gas Pressure sensor you want to connect to. If needed check the sensor's Bluetooth ID which is located on the sensor's bottom label. If the sensor was not paired yet Coach will force you to pair the sensor first via Windows Settings.
- When the connection is established the Bluetooth symbol appears in the top-left corner of the sensor's display and the icon of the sensor appears showing the measured values.
- Now you are ready to use the Gas Pressure sensor for your measurement.

## Collecting data via the USB connection

For computers (Windows and Mac) the Gas Pressure sensor can also be used as USB sensor. When using this connection, the sensor can measure with a higher sampling frequency of up to 1000 Hz.

- Turn the Gas Pressure sensor on.
- Use the provided USB cable to connect the sensor to a USB port.
- Start the Coach 7 or Coach 7 Lite program.
- Select the Dashboard Activity 'Measurement with Wireless sensors'.
- The connected USB sensor should be detected automatically, and its icon appears on the first empty sensor position in the Wireless sensors panel.
- When the connection is established the USB symbol appears in the top-left corner of the sensor's display and the sensor icon shows measured data.
- Now you are ready to use the Gas Pressure sensor for your measurement.

## Charging a battery

An internal rechargeable battery (Li-Poly 3.7 V, 700 mAh) powers the sensor. The battery symbol located in the top-right corner of the sensor's display shows the battery level. When the battery level becomes critical, the battery gauge shows an empty battery. Use the provided cable to connect the sensor to a USB port for charging. A fully discharged battery requires up to 2 hours of charge time to become fully charged again. To prolong battery life, automatic power down turns the sensor off after 5 minutes of inactivity.

To replace the battery, use **only** the approved rechargeable batteries provided by CMA.

## Suggested experiments

This Gas Pressure sensor can be used in various experiments such as:

- Investigating the relationship between pressure and volume, Boyle's law,
- study the effect of temperature on gas pressure, Gay-Lussac's law,
- measurement of vapour pressure of liquids,
- measurements of reaction rates as a gas is produced in a chemical reaction.

## Pressure Units

Pressure can be measured in many different units.

Some equivalent values for 1 atmosphere are:

1 atmosphere = 101.325 kPa

1 atmosphere = 760 mm Hg

1 atmosphere = 29.92 in. of Hg (at 0°C)

1 atmosphere = 14.70 psi

1 atmosphere = 1013 millibar

## Technical Specifications

|                                       |   |
|---------------------------------------|---|
| <i>Sensor kind</i>                    | Digital, on-sensor digital conversion, 14-bit resolution  |
| <i>Measuring range</i>                | - 100 .. 300 kPa, relative to atmospheric pressure  |
| <i>Resolution</i>                     | Typical 0.06 kPa  |
| <i>Accuracy</i>                       | Typical $\pm 1.0$ % of full range   |
| <i>Connector</i>                      | Pressure port female Luer-lock  |
| <i>Usage</i>                          | Only for non-corrosive gasses, non-ionic working gasses such as air, dry gases and the like   |
| <i>Respond time</i>                   | 90% of applied pressure value: typical, 0.2 ms (before digital condition and conversion)  |
| <i>Maximal sampling rate</i>          | 100 Hz  |
| <i>Display</i>                        | OLED 0.96" (128*64 px)  |
| <i>Battery</i>                        | Li-Poly Rechargeable Battery (3,7 V 700 mAh)  |
| <i>Battery life after full charge</i> | Approximately 13 hours<br>Battery life varies by use, configuration, temperature, and many other factors; actual results will vary. |
| <i>Connection</i>                     | Bluetooth 5, Low Energy (Mac, Android, iOS)<br>Bluetooth 2.1, Classic (Windows)<br>USB 2.0 (type C)                                 |
| <i>Bluetooth ID</i>                   | W24PRES-xxx   |
| <i>Sensor dimensions</i>              | Housing: 80 x 50 x 25 mm<br>Length of female Luer-lock connector: 6.2 mm  |

## Warranty

The Conductivity sensor W04 is warranted to be free from defects in materials and workmanship for a period of 3 years from the date of purchase provided that it has been used under normal laboratory conditions. This warranty does not apply if the sensor has been damaged by accident or misuse.

The sensor battery is consumable and is warranted to be free from defects in materials and workmanship for a period of 12 months from the date of purchase.

Discard batteries according to local regulations.



**Note:** This product is to be used for educational purposes only.  
It is not intended for industrial, medical, research, or commercial applications.

Rev. 01.09.2025