

Franck-Hertz Tube with Ne Filling on Base 1000912

Instruction sheet

10/15 ALF



- 1 BNC connector
- 2 Screening cylinder with observation window
- 3 Franck-Hertz tube
- 4 Base with connector sockets

1. Safety instructions

- Do not subject the tube to any mechanical stress. Do not put kinks in any connecting leads. There is always a risk that glass can break and cause injury.

2. Description

The Franck-Hertz tube is a tetrode with an indirectly heated barium oxide cathode K, a mesh-type control grid G, a mesh-type anode A, and a collector electrode E (see Fig. 1). The electrodes are in a plane-parallel configuration. The distance between the control grid and the anode grid is about 5 mm, and the distances between the cathode and the control grid and between the anode

and the collector electrode are both about 2 mm. The tube is supplied already filled with neon gas at a pressure chosen to give an optimum characteristic curve, which is in the region of several hPa.

The connecting sockets for the heater, control grid and anode grid voltages are on the base of the instrument. The collector current is taken off through the BNC socket at the top end of the screening cylinder. An internal 10 k Ω limiting resistor is permanently built in between the connector sockets for the accelerator (control grid) voltage and the anode voltage. This protects the tube in case there is a spark discharge caused by applying too high a voltage. The voltage loss in this resistor when making measurements is negligible, as the anode current in the tube is smaller than 5 pA. (Thus the voltage loss in the protecting resistor is 0.05 V.)

3. Technical data	
Filament voltage:	4 – 12 V
Control voltage:	9 V
Accelerating voltage:	max. 80 V
Countervoltage:	1.2 – 10 V
Tube:	130 x 26 mm dia. approx.
Base with connector sockets:	190x115x115 mm ³ approx.
Weight:	450 g approx.

4. Basic principles

In the Franck-Hertz experiment neon atoms are excited by inelastic collision with electrons. The excited atoms emit visible light that can be viewed directly. Thus it is possible to detect zones where the light and therefore the excitation is more intense. The distribution of such zones between the cathode and the grid depends on the difference in potential between the two:

Electrons are emitted from the cathode and are accelerated by a voltage U towards the grid. Having passed through the grid they reach the target and thus contribute to a target current I if their kinetic energy is sufficient to overcome a decelerating voltage between the grid and the target.

The $I(U)$ -characteristic (see Fig. 3) has a similar pattern to the original Franck-Hertz experiment using mercury gas but this time the intervals between minima where the current falls to almost zero for a specific voltage $U = U_1$ corresponding to the electrons reaching sufficient kinetic energy to excite a neon atom by inelastic collision just before reaching the grid are about 19 V. Simultaneously it is possible to observe a faint orange light close to the grid since the energy transition to the base state of a neon atom results in the emission of such light. The zone of illumination moves towards the cathode as the voltage U increases and the target current I rises once more.

For a higher voltage $U = U_2$ the target current also drops drastically and it is possible to see two zones of illumination. The electrons can in this case retain enough energy after an initial collision to excite a second neon atom.

As the voltages are further increased, other minima in the target current along with further zones of illumination can be observed.

The $I(U)$ -characteristic exhibits various maxima and minima and the interval between the minima is about $\Delta U = 19$ V. This corresponds to excitation energy of the 3p energy level of a neon atom (see Fig. 4) so that it is highly likely that this level is being excited. Excitation of the 3s-level can-

not be neglected entirely and gives rise to some fine detail in the structure of the $I(U)$ -characteristic.

The zones of illumination are zones of greater excitation and correspond to drops in voltage in the $I(U)$ -characteristic. One more zone of illumination is created every time U is increased by about 19 V.

Note

The first minimum is not at 19 V itself but is shifted by an amount corresponding to the so-called contact voltage between the cathode and grid.

The emission lines in the neon spectrum can easily be observed and measured using a spectroscope (1003184) when the maximum voltage U is used.

5. Operation

The following equipment is also required to complete the experiment:

1 Power supply unit for F/H experiment @230 V	1012819
or	
1 Power supply unit for F/H experiment @115 V	1012818
1 Analogue oscilloscope, 2x 30 MHz	1002727
1 HF Patch cord, 1 m	1002746
2 HF Patch cords, BNC/4 mm plug	1002748
Safety leads for experiments	1002843

- Start with the voltage supply unit switched off, and with all the voltage setting knobs fully to the left.
- Connect up the experiment as shown in Fig. 2.
- Turn on the equipment. It will start in ramp mode.
- Set up the oscilloscope in XY mode with $x = 1$ V/div and $y = 2$ V/div.
- Gradually increase the heater voltage till the filament starts to faintly glow red. Then wait 30 seconds till it reaches its operating temperature.
- Set the accelerating voltage to 80 V and select a grid voltage of 9 V.

The ideal heater voltage should be between 4 and 12 V. This differs from tube to tube due to manufacturing tolerances.

- Gradually increase the heater voltage until an orange glow appears between the cathode and the grid. Then turn down the heater voltage till the glow disappears and only the filament is glowing.

- Gradually increase the decelerating voltage until the measured curve (of signal against accelerating voltage) is near horizontal.
- Increase the gain till the maxima of the Franck-Hertz curve can be seen on the oscilloscope screen.

6. Disposal

- The packaging should be disposed of at local recycling points.
- Should you need to dispose of the equipment itself, never throw it away in normal domestic waste. Local regulations for the disposal of electrical equipment will apply.





Fig. 1 Schematic of set up for measuring the Franck-Hertz curve for neon (K cathode, G control grid, A anode, E collector electrode)




Fig. 2 Experiment set-up - Franck-Hertz tube filled with neon




Fig. 3 Target current I as a function of the accelerating voltage U




Fig. 4 Energy levels in neon atoms